AMETEK SI · PRINCETON APPLIED RESEARCH

QA-A9M-AL(M)-25 QCM-resonatorer

QCM accessory

QA-A9M-AL(M)-25 by Princeton Applied Research. Qcm accessory.

Coating material	Pack count	Instrument
Aluminum [1]	**25** [1]	**QCM922A** [1]

Specifications

Coating material	Aluminum	[1]
Pack count	25	[1]
Instrument	QCM922A	[1]
Resonator type	9 MHz AT-cut quartz crystal	[2]
Electrode area	Ø5 mm (19.6 mm²)	[2]
Electrode film	300 nm electrode material sputtered onto a Ti film	[2]

Sources

[1] **QCM922A Quartz Crystal Microbalance and EQCM**
AMETEK Scientific Instruments · product-page · Options – Resonators table; QCM Family; Options – Resonators (QCM922A page) · checked 2026-10-16
https://www.ameteksi.com/products/electrochemical-accessories/qcm922a-quartz-crystal-microbalance-eqcm

[2] **QCM922A Quartz Crystal Microbalance (brochure, Rev 1219)**
AMETEK Scientific Instruments · brochure · p. 2, Specification – 9MHz AT-Cut Quartz Crystal Resonator QA-A9M-series; p. 2, Specification – QA-A9M-series · checked 2026-10-16
https://www.ameteksi.com/-/media/ameteksi/download_links/documentations/qcm922/qcm922a_brochure.pdf?revision...

Every value is taken from the manufacturer's own document; the number after the value refers to the source list. Always check against the latest datasheet before ordering.

The data comes from the manufacturer's published documents. Manufacturers may change specifications; the current configuration and performance are confirmed in the quotation.